

Attaching the Thoracic Extension to a Pacific™ Adjustable Collar

1. Remove the anterior lower liner from the Pacific™ Adjustable Collar.
2. Remove existing Velcro® coins from the areas around the (3) small holes on the underside of the anterior lower shell.
3. Insert one of the supplied bushing through any of the three holes so it is seated underneath the shell.
4. On the front side of the anterior lower shell, align the anterior yoke attachment hole with the bushing.
5. Secure the anterior yoke in place using the supplied thumbscrew. Repeat steps 3 – 5 until all three anterior yoke attachment points are secure.
6. Check that the anterior yoke is secured firmly to the Pacific™ Adjustable Collar.
7. Apply new Velcro® Coins supplied to the areas they were removed from in step 2.
8. Reattach anterior lower liner to the Pacific™ Adjustable Collar.
9. Repeat steps 1-8 to apply back plate to posterior neck section.

Application Instructions

1. Adjust Pacific™ Adjustable Collar per instructions.
2. Adjust height of the anterior portion of the Thoracic Extension to your desired position and secure using the supplied thumbscrews. A hex wrench may be used for additional torque if necessary.
3. Adjust height of the posterior section of the Thoracic Extension to your desired position and secure using the supplied thumbscrews. If your patient will be lying on his/her back for extended periods of time, we suggest replacing the thumbscrews with the provided pan head screws.
4. If needed, contour the Para Spinal and Sternal Bars to fit patient.
Caution: Isolate bending force below attachment points, failure to do so may damage brace.
5. Secure to the patient with waist straps following appropriate spinal precautions. Only trained personal should attempt logroll application
6. To shorten the length of the straps, remove the double-sided hook from the ends, cut excess length, and re-apply.

Attaching the Thoracic Extension to a Pacific™ Adjustable Collar

1. Remove the anterior lower liner from the Pacific™ Adjustable Collar.
2. Remove existing Velcro® coins from the areas around the (3) small holes on the underside of the anterior lower shell.
3. Insert one of the supplied bushing through any of the three holes so it is seated underneath the shell.
4. On the front side of the anterior lower shell, align the anterior yoke attachment hole with the bushing.
5. Secure the anterior yoke in place using the supplied thumbscrew. Repeat steps 3 – 5 until all three anterior yoke attachment points are secure.
6. Check that the anterior yoke is secured firmly to the Pacific™ Adjustable Collar.
7. Apply new Velcro® Coins supplied to the areas they were removed from in step 2.
8. Reattach anterior lower liner to the Pacific™ Adjustable Collar.
9. Repeat steps 1-8 to apply back plate to posterior neck section.

Application Instructions

1. Adjust Pacific™ Adjustable Collar per instructions.
2. Adjust height of the anterior portion of the Thoracic Extension to your desired position and secure using the supplied thumbscrews. A hex wrench may be used for additional torque if necessary.
3. Adjust height of the posterior section of the Thoracic Extension to your desired position and secure using the supplied thumbscrews. If your patient will be lying on his/her back for extended periods of time, we suggest replacing the thumbscrews with the provided pan head screws.
4. If needed, contour the Para Spinal and Sternal Bars to fit patient.
Caution: Isolate bending force below attachment points, failure to do so may damage brace.
5. Secure to the patient with waist straps following appropriate spinal precautions. Only trained personal should attempt logroll application
6. To shorten the length of the straps, remove the double-sided hook from the ends, cut excess length, and re-apply.

Attaching the Thoracic Extension to a Pacific™ Adjustable Collar

1. Remove the anterior lower liner from the Pacific™ Adjustable Collar.
2. Remove existing Velcro® coins from the areas around the (3) small holes on the underside of the anterior lower shell.
3. Insert one of the supplied bushing through any of the three holes so it is seated underneath the shell.
4. On the front side of the anterior lower shell, align the anterior yoke attachment hole with the bushing.
5. Secure the anterior yoke in place using the supplied thumbscrew. Repeat steps 3 – 5 until all three anterior yoke attachment points are secure.
6. Check that the anterior yoke is secured firmly to the Pacific™ Adjustable Collar.
7. Apply new Velcro® Coins supplied to the areas they were removed from in step 2.
8. Reattach anterior lower liner to the Pacific™ Adjustable Collar.
9. Repeat steps 1-8 to apply back plate to posterior neck section.

Application Instructions

1. Adjust Pacific™ Adjustable Collar per instructions.
2. Adjust height of the anterior portion of the Thoracic Extension to your desired position and secure using the supplied thumbscrews. A hex wrench may be used for additional torque if necessary.
3. Adjust height of the posterior section of the Thoracic Extension to your desired position and secure using the supplied thumbscrews. If your patient will be lying on his/her back for extended periods of time, we suggest replacing the thumbscrews with the provided pan head screws.
4. If needed, contour the Para Spinal and Sternal Bars to fit patient.
Caution: Isolate bending force below attachment points, failure to do so may damage brace.
5. Secure to the patient with waist straps following appropriate spinal precautions. Only trained personal should attempt logroll application
6. To shorten the length of the straps, remove the double-sided hook from the ends, cut excess length, and re-apply.

PACIFIC™

Adjustable Collar with Thoracic Extension



PACIFIC™

Adjustable Collar with Thoracic Extension



PACIFIC™

Adjustable Collar with Thoracic Extension



Thoracic Extension Attachment/Application Instructions



635 Executive Dr | Troy, MI 48083
P: 800-521-2192 | 248-588-7480
BeckerOrthopedic.com | f t in

EC REP

Acorn Regulatory Consultancy
Services Limited
Knockmorris Cahir Co. Tipperary Ireland,
Postcode: E21 R766
P 012 4626 8456
F 012 4626 8648

©2021 Becker Orthopedic Appliance Co.
All rights reserved.
Revision 05/01/21



Thoracic Extension Attachment/Application Instructions



635 Executive Dr | Troy, MI 48083
P: 800-521-2192 | 248-588-7480
BeckerOrthopedic.com | f t in

EC REP

Acorn Regulatory Consultancy
Services Limited
Knockmorris Cahir Co. Tipperary Ireland,
Postcode: E21 R766
P 012 4626 8456
F 012 4626 8648

©2021 Becker Orthopedic Appliance Co.
All rights reserved.
Revision 05/01/21



Thoracic Extension Attachment/Application Instructions



635 Executive Dr | Troy, MI 48083
P: 800-521-2192 | 248-588-7480
BeckerOrthopedic.com | f t in

EC REP

Acorn Regulatory Consultancy
Services Limited
Knockmorris Cahir Co. Tipperary Ireland,
Postcode: E21 R766
P 012 4626 8456
F 012 4626 8648

©2021 Becker Orthopedic Appliance Co.
All rights reserved.
Revision 05/01/21

